



माडी नगरपालिका

नगर कार्यपालिकाको कार्यालय

बसन्तपुर, चितवन

बागमती प्रदेश, नेपाल

२०७३



प.सं. २०८३/८४

उच.नं. ३४



मिति: २०८३/०६/२१

बिषय : बजार दररेट उपलब्ध गराइदिने सम्बन्धमा ।

(तेश्रो पटक प्रकाशित सूचना मिति: २०८३/०६/२१)

श्री सबै फर्म तथा कम्पनीहरु,

.....।

प्रस्तुत बिषयमा यस कार्यालय अन्तर्गतको बघौडा अस्पतालमा पोस्टमार्टम सेवाका लागि यसैसाथ संलग्न विवरण तथा स्पेशिफिकेशन बमोजिमको Mortuary (Chamber-2 Body) Fridge को बजार दररेट आवश्यक परेको हुँदा सूचना प्रकाशन भएको मितिले ७ (सात) दिन (मिति २०८३/०६/२७ गते) भित्र उपलब्ध गराइदिनुहुन यो सूचना प्रकाशन गरिएको छ ।

दररेट माडी नगरपालिकाको कार्यालयमा सोझै वा E-mail मार्फत पेश गर्न सकिनेछ ।

E-mail: madihealthunit@gmail.com

.....


जनक थापा

प्रमुख प्रशासकीय अधिकृत

प्रमुख प्रशासकीय अधिकृत

Technical Specification of Mortuary Chamber-2 Body

S.N.	Purchaser's Specifications	Bidders Remarks	
		Yes/No	Page No. in Catalogue
	Mortuary Chamber-2 Body		
	Manufacturer:		
	Brand:		
	Type / Model:		
	Country of Origin:		
1	Description of Function		
1.1	Mortuary Chamber is used for preservation of dead bodies under controlled low temperature to prevent decomposition.		
2	Operational Requirement		
2.1	Two Body Mortuary Chamber unit having temperature range of -2°C to +6°C.		
3	System Configuration		
3.1	Mortuary Chamber with CFC free refrigeration System.		
4	Technical Specification		
4.1	Double walled chamber with inner body and outer body made up of 304 quality SS.		
4.2	Insulation: High grade Poly Urethane with double walled Insulation of at least 60mm.		
4.3	Doors made up of corrosion resistance Stainless steel (304 or 316)		
4.4	Stretcher made up of Stainless steel 304 or 316 and sliding.		
4.5	Microprocessor based system with LCD/LED display with vapour proof light.		
4.6	Each chamber shall be independently temperature controlled.		
4.7	Compressor : CFC Free		
4.8	System should have full unit easily movable on castor wheels with lockable braking.		
4.9	It should have various audio and visual alarms for various failures.		
4.10	Locking System: Standard key locks.		
5	Accessories, Spares and Consumables		
5.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer		
6	Operating Environment		
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.		
7	Standards and Safety Requirements		


 Biomedical Engineer
 NEC No. - 83941